




NOVA TESTINGS PTE LTD

280 Woodlands Industrial Park E5 #02-41 Harvest @ Woodlands Singapore 757322 Tel: +65 6561 1002/6 Fax: +65 6561 2003

E-mail: sales@novatestings.com Website: www.novatestings.com

Type SLB load cell



Description


Flintec loadcells are designed to meet the most stringent accuracy requirements. Certifications have been obtained from Weights & Measures Authorities ,worldwide.

SB4 loadcells areavailable in thecapacities 5 kN to 100 kN (510kgto10197kg) and include Accuracy Classifications GP, C1, C3, C3 MI 7.5, C4 and C4 MI7.5 according to OIML R 60; NTEP n max=5000.

They offer total stainless steel construction and complete hermetic sealing, making them suitable for use in the toughest industrial environments.

The unique "blind" loading hole combined with the available Flintec loading hardware provides an excellent price-performance ratio.

It allows very low profile platform design and offers advantages in all kinds of weighing applications. The Flintec calibration technique (inmV/V/) eliminates time consuming corner calibration in multiple load cell systems. The SLB is available for use in hazardous areas zone 1, 2 (gas) and 21, 22 (dust) according to EExia IIC T6...T4 T130°CATEX.




All dimensions in mm. Dimensions and specifications are subject to change without notice.

Mounting bolts for 200 lb to 2500 lb: M12 8.6 / torque 90 Nm; for 5000 lb: M12 10.9 / torque 120 Nm. Torque values assume oiled threads.

* Unified thread 1/2-20 UNF is available (type designation SLB-xx lb-Cx-CU).

SLB Specifications

| lb | 200 / 500 / 1000 | 200 / 500 / 1000 / 2500 / 5000 | |
|-------------------|---|--|--|
| kg | 91 / 227 / 454 / | 91 / 227 / 454 / 1134 / 2268 | |
| mV/V | 2 ± 0. | 2 ± 0.1% | |
| %RO | ± 0.05(| ± 0.05(± 0.005) | |
| | (GP) | СЗ | |
| | n.a. | 3000 | |
| ĺ | n.a. | E _{max} /11500 | |
| %RO | ± 0.940 | ± 0.020 | |
| %RO | ± 0.060 | ± 0.024 | |
| %Ro/°C | ± 0.0040 | ± 0.0011 | |
| %/°C | ± 0.0020 | ± 0.0011 | |
| V | 51 | 515 | |
| %RO | ± 1 | ± 1.0 | |
| | 1106 ± | 1106 ± 5 | |
| | 1000 ± | 1000 ± 1 | |
| M | 500 | 5000 | |
| °C. | –1 0+ | –1 0 +40 | |
| °C | –2 0+ | -2 0+65 | |
| %E _{max} | 200 | 200 | |
| %E _{nax} | 300 | 300 | |
| %E _{max} | 100 | 100 | |
| | stainless steel 17-4 PH (1.4548) | | |
| | potted | | |
| | IP 67 | | |
| | kg mV/V %RO %RO %RO %RO/°C V %RO M °C °C %E _{max} %E _{reax} | kg 91/227/454/ mV/V 2±0. %RO ±0.05(GP) n.a. n.a. %RO ±0.040 %RO ±0.060 %RO/*C ±0.0020 V 51: %RO ±10.0020 M 500: M 500: **C -10+ **C -20+ %Enax 300 %Enax 1000 stainless steel 17- pottee | |



Wiring

- The load cell is provided with a shielded, 4 conductor cable.
- · Cable length: 3 m
- Cable diameter: 5 mm
- The shield is connected to the load cell body.

Note:

Maximum errors (+/-) are expressed as a percentage of rated load (when measured at 10Vdc excitation). The linearity and hystersis is the maximum deviation from a straight line drawn between the no load and the rated load outputs for both increasing and decreasing loads. Accuracy classes apply when active range (live weight) is within 20 to 90% of rated load.

